

Schmizo Shukoff-Flasks


vacuum jacketed, for the determination of the cooling curve of cocoa butter according to IOCCC, calibration protocol included



Schmizo Shukoff-Flasks

(Laboratory Glassware page 152)

- Art. 25.100.01: **NS 19/26**
- Art. 25.100.12: **NS 24/29**
- **Serial-Number**
- **Capacity Marks**
- **Calibration Protocol**



Standard Labeling

- **Ground Size**
- **Serial-Number**
- **Schmizo-Logo**
- **2 Capacity Marks**
(12.5/25g Soybeanoil)



Calibration Protocol

Shukoff Flask no: 624

Printing Date: 29.05.2018

Measured Values (°C / min)						
Date	00:00	05:00	10:00	15:00	20:00	25:00
x1	01.01.2016	39.9	32.7	27.0	22.4	18.7
x2	01.01.2016	39.8	32.7	27.0	22.4	18.7
x3	01.01.2016	40.2	32.8	27.0	22.5	18.8
x4	01.01.2016	0.0	0.0	0.0	0.0	0.0
x5	01.01.2016	0.0	0.0	0.0	0.0	0.0
x6	01.01.2016	0.0	0.0	0.0	0.0	0.0
x7	01.01.2016	0.0	0.0	0.0	0.0	0.0
x8	01.01.2016	0.0	0.0	0.0	0.0	0.0

Default Values (°C / min)					
	00:00	05:00	10:00	15:00	20:00
Min	39.7	32.5	26.9	22.1	18.2
Average	40.0	32.8	27.2	22.4	18.5
Max	40.3	33.1	27.5	22.7	18.8

Vacuum: 0.01 min Hg
 0.013 mbar

Temperature in °C

